

- Program by the person holding the position listed in Column 2 of Schedule 1, and has successfully completed training approved by the Environment Protection Authority; and
2. the exposure is limited to a person who has been classified by the Registered Nurse giving the authorisation as triage category 3, 4 or 5; and
  3. limited to the authorisation by the Registered Nurse for the purpose of plain radiography; limited to the items and regions specified in Schedule 2 below; and
  4. in addition to the requirements of Regulation 106(2) of the *Radiation Protection and Control Regulations 2022*, the authorisation referred to in condition 3 must:
    - (a) indicate that the exposure is authorised as part of the NIXR Program; and
    - (b) indicate that the person to be examined has been classified as triage category 3, 4 or 5; and
  5. the person holding or acting in the position listed in Column 2 of The Schedule must, when directed by the Environment Protection Authority, provide a written report on diagnostic radiography authorised by the Registered Nurses per the direction. The report must contain statistics regarding examinations authorised by the Registered Nurse(s), details of any radiation safety issues associated with examinations, or any other matter as required by the direction.

## SCHEDULE 1

Column 1	Column 2
Royal Adelaide Hospital	Site Director, Emergency Department RAH
Lyell McEwin Hospital	Director, Emergency Department, Lyell McEwin Hospital
Flinders Medical Centre	Deputy Network Clinical Director—FMC Emergency Department
The Queen Elizabeth Hospital	Director, Emergency Department
Noarlunga Health Services	Clinical Director Emergency Department, Noarlunga Health Service
Women's and Children's Hospital	Medical Unit Head, Paediatric Emergency Department

## SCHEDULE 2

Item	Region
57509	Hand, wrist, forearm, elbow or humerus (R)
57515	Hand and wrist, or hand, wrist and forearm, or forearm and elbow, or elbow and humerus (R)
57521	Foot, ankle, leg, knee or femur (R)
57527	Foot and ankle, or ankle and leg, or leg and knee, or knee and femur (R)
57703	Shoulder or scapula (R)
57709	Clavicle (R)
57712	Hip Joint (R)
57715	Pelvic Girdle (R)
57721	Femur, internal fixation or neck or intertrochanteric (perthrochanteric) fracture (R)

The exemption shall take effect commencing on the date of publication of this Notice in the *Government Gazette*.

Dated: 11 January 2024

D. BELLIFEMINE  
Delegate of the Minister for Climate, Environment and Water

## RADIATION PROTECTION AND CONTROL ACT 2021

## SECTION 76

## Notice RPC202401-08

Pursuant to Section 76 of the *Radiation Protection and Control Act 2021* (the Act), I, Keith Baldry, Director Science and Systems, Environment Protection Authority, Delegate of the Minister for Climate, Environment and Water, exempt persons who use or operate a portable X-ray fluorescence (pXRF) apparatus from any requirement in the *Code of practice for protection against ionizing radiation emitted from X-ray analysis equipment* (1984) published by ARPANSA, subject to the following condition:

1. The registered owner of the Portable X-ray fluorescence (pXRF) apparatus ensures that the following tables from Section 4 of the *Guideline for pXRF Apparatus 2023* as published by the Department, are complied with—
  - (a) Table 4.1 Apparatus requirements for Portable X-ray fluorescence (pXRF) apparatus
  - (b) Table 4.2 Operational requirements for Portable X-ray fluorescence (pXRF) apparatus

Dated: 11 January 2024

K. BALDRY  
Delegate of the Minister for Climate, Environment and Water